

Declaration of Conformity

Reference:	2245
-------------------	------

(Note 1)

I / We (Dealer): Brüel & Kjær Sound & Vibration Measurement A/S

Of Skodsborgvej 307, DK-2850 Nærum, Denmark

declare under my/our sole responsibility that the product(s)

Brüel & Kjær Hand-held Analyzer, model B&K 2245

to which this declaration relates, is/are in conformity with the TRA Telecommunication Apparatus – Type Approval Regulations and the following technical Standards.

Article 1 : Safety and Health : EN 61010-1: 2010; EN 62209-2:2010.

Article 2 : EMC : EN 61326-1:2013; EN 301 489-1 V2.2.0; EN 301 489-17 V3.2.0; EN 301 489-19 V2.1.0.


Article 3 :

Frequency	Power	RF Spectrum Efficiency Standard
2400-2483.5 MHz	WLAN: 19.86dBm; BT/BLE: 14.81dBm	EN 300 328 V2.1.1 EN 300 328 V2.1.1
GPS/GLONASS	N/A (Receiver)	EN 303 413 V1.1.1

and that I have examined the technical basis for this declaration which is based on test reports and /or certificates issued by

TUV Rheinland, CETECOM GmbH, International Certification Corp.

Supporting documentation relevant to this declaration has been compiled under the reference number given above and will be made available to TRA on request.

Signature  **Date:** 2019.08.12

Lars Thorndahl Kroman, Technology Manager

Name & Title

NOTES:

- 1) *The reference is a unique identifier for the declaration of conformity (DoC) allocated by the manufacturer / supplier himself for the purpose of cross reference in the registration application.*